

**FLUKE** ®

Calibration



## 5730A High Performance Multifunction Calibrator Extended specifications

## General Specifications

**Warm-Up Time** ..... Twice the time since last warmed up, to a maximum of 30 minutes.

**System Installation** ..... Rack mount kits available.

**Standard Interfaces** ..... IEEE-488, RS-232, USB 2.0 device, Ethernet, 5725A, 52120A, phase lock in (BNC), phase reference out (BNC).

### Temperature Performance

Operating ..... 0 °C to 50 °C

Calibration ..... 15 °C to 35 °C

Storage ..... -40 °C to 75 °C

### Relative Humidity

Operating ..... <80 % to 30 °C, <70 % to 40 °C, <40 % to 50 °C

Storage ..... <95 %, non-condensing. A power stabilization period of four days may be required after extended storage at high temperature and humidity.

**Safety** ..... IEC 61010-1: Overvoltage Category II, Pollution Degree 2

**Operating Altitude** ..... 2000 m maximum

**Guard Isolation** ..... 20 V

### Electromagnetic Compatibility (EMC)

IEC 61326-1 (Controlled EM environment) .... IEC 61326-2-1; CISPR 11: Group 1, Class A

*Group 1: Equipment has intentionally generated and/or uses conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.*

*Class A equipment is equipment suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.*

*Emissions which exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object. The equipment may not meet the immunity requirements of 61326-1 when test leads and/or test probes are connected.*

USA (FCC) ..... 47 CFR 15 subpart B, this product is considered an exempt device per clause 15.103

Korea (KCC) ..... Class A Equipment (Industrial Broadcasting & Communication Equipment)

*This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.*

### Line Power

#### Line Voltage

5730A ..... 100 V-120 V, 220 V- 240 V ±10 %

5725A ..... 100 V, 110 V, 115 V, 120 V, 200 V, 220 V, 230 V, 240 V, ±10 %

Line Frequency ..... 47 Hz-63 Hz

#### Maximum Power

5730A ..... 300 VA

5725A ..... 750 VA

### Weight

5730A ..... 27 kg (62 lb)

5725A ..... 32 kg (70 lb)

### Size

#### 5730A

Height ..... 17.8 cm (7 in), standard rack increment, plus 1.5 cm (0.6 in) for feet


Width ..... 43.2 cm (17 in), standard rack width

Depth ..... 64.8 cm (25.5 in), overall; 59.4 cm (23.4 in), rack depth

#### 5725A

Height ..... 13.3 cm (5.25 in)

Width and Depth ..... Both units project 5.1 cm (2 in) from rack front.



hhp002.eps

**Figure 1-1. Product Dimensions****Artifact Calibration Standards Requirements**

The following external standards are necessary to calibrate the 5730A to the listed specification. Each external standard used must have an uncertainty equal to or less than the listed uncertainty limit.

<b>Fluke Standard</b>	<b>Traceable Quantity</b>	<b>Nominal Value</b>	<b>Uncertainty Limit</b>	<b>5730A Specifications Susceptible to Uncertainty Limit</b>
732B	Voltage	10 V	1.5 ppm	dc volts, ac volts, dc current, ac current
742A-1	Resistance	1 Ω	10 ppm	1 Ω, 1.9 Ω
742A-10k	Resistance	10 kΩ	2 ppm	ac current, dc current 10 Ω to 100 MΩ

## Electrical Specifications

The product specifications describe the Absolute Instrumental Uncertainty of the Product. The product specifications include stability, temperature, and humidity; within specified limits, linearity, line and load regulation, and the reference standard measurement uncertainty. The product specifications are provided at a 99 %, k=2.58, normally distributed and a 95 %, k=2, normally distributed level of confidence. Fluke Calibration guarantees product performance to the 99 % level of confidence.

The relative specifications are provided for enhanced accuracy applications. The specifications apply when range constants are adjusted (see "Range Calibration"). To calculate an enhanced absolute specification from the relative accuracy specification, it is necessary to combine the uncertainty of your external standards with the pertinent relative specifications.

Specifications are valid after allowing a warm-up period of 30 minutes, or twice the time the Product has been turned off.  
DC Voltage Specifications

### 5730A DC Voltage Specifications

Range	Resolution	Absolute / $\pm 5^{\circ}\text{C}$ from calibration temperature				Relative $\pm 1^{\circ}\text{C}$	
		24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days
		$\pm(\text{ppm output}^{[1]} + \mu\text{V})$					
<b>99 % Confidence Level</b>							
220 mV	10 nV	5 + 0.5	7 + 0.5	8 + 0.5	9 + 0.5	2 + 0.4	2.5 + 0.4
2.2 V	100 nV	3.5 + 0.8	4 + 0.8	4.5 + 0.8	6 + 0.8	2 + 0.8	2.5 + 0.8
11 V	1 $\mu\text{V}$	2.5 + 3	3 + 3	3.5 + 3	4 + 3	1 + 3	1.5 + 3
22 V	1 $\mu\text{V}$	2.5 + 5	3 + 5	3.5 + 5	4 + 5	1 + 5	1.5 + 5
220 V	10 $\mu\text{V}$	3.5 + 50	4 + 50	5 + 50	6 + 50	2 + 50	2.5 + 50
1100 V	100 $\mu\text{V}$	5 + 500	6 + 500	7 + 500	8 + 500	2.5 + 400	3 + 400
<b>95 % Confidence Level</b>							
220 mV	10 nV	4 + 0.4	6 + 0.4	6.5 + 0.4	7.5 + 0.4	1.6 + 0.4	2 + 0.4
2.2 V	100 nV	3 + 0.7	3.5 + 0.7	4 + 0.7	5 + 0.7	1.6 + 0.7	2 + 0.7
11 V	1 $\mu\text{V}$	2 + 2.5	2.5 + 2.5	3 + 2.5	3.5 + 2.5	0.8 + 2.5	1.2 + 2.5
22 V	1 $\mu\text{V}$	2 + 4	2.5 + 4	3 + 4	3.5 + 4	0.8 + 4	1.2 + 4
220 V	10 $\mu\text{V}$	3 + 40	3.5 + 40	4 + 40	5 + 40	1.6 + 40	2 + 40
1100 V	100 $\mu\text{V}$	4 + 400	4.5 + 400	6 + 400	6.5 + 400	2 + 400	2.4 + 400

Notes:

Perform the DC Zero calibration every 30 days. In addition, perform the DC Zero calibration after powering up the unit the first time after unpacking following a shipment or if exposed to an environmental change of greater than  $5^{\circ}\text{C}$ .

- For radiated EMI fields >400 MHz and <500 MHz, add 1 ppm.

### DC Voltage Secondary Performance Specifications and Operating Characteristics

Range	Stability <sup>[1]</sup> $\pm 1^{\circ}\text{C}$ 24 Hours	Temperature Coefficient Adder <sup>[2]</sup>		Linearity $\pm 1^{\circ}\text{C}$	Noise	
		10 - 40 °C	0 - 10 °C and 40 - 50 °C		Bandwidth 0.1 - 10 Hz pk-pk	Bandwidth 10 - 10 kHz RMS
		$\pm(\text{ppm output} + \mu\text{V})$	$\pm(\text{ppm output} + \mu\text{V}) / ^{\circ}\text{C}$		$\pm(\text{ppm output} + \mu\text{V})$	$\mu\text{V}$
220 mV	0.3 + 0.3	0.4 + 0.1	1.5 + 0.5	1 + 0.2	0.15 + 0.1	5
2.2 V	0.3 + 1	0.3 + 0.1	1.5 + 2	1 + 0.6	0.15 + 0.4	15
11 V	0.3 + 2.5	0.15 + 0.2	1 + 1.5	0.3 + 2	0.15 + 2	50
22 V	0.4 + 5	0.2 + 0.4	1.5 + 3	0.3 + 4	0.15 + 4	50
220 V	0.5 + 40	0.3 + 5	1.5 + 40	1 + 40	0.15 + 60	150
1100 V	0.5 + 200	0.5 + 10	3 + 200	1 + 200	0.15 + 300	500

Notes:

- Stability specifications are included in the absolute specification values in the primary specification tables.
- Temperature coefficient is an adder to accuracy specifications that does *not* apply unless operating more than  $\pm 5^{\circ}\text{C}$  from calibration temperature.

**Minimum Output** ..... 0 V for all ranges, except 100 V for 1100 V range

**Maximum Load** ..... 50 mA for 2.2 V through 220 V ranges; 20 mA for 1100 V range;  
50  $\Omega$  output impedance on 220 mV range; all ranges <1000 pF,  
 $>25 \Omega$

**Load Regulation** .....  $<(0.2 \text{ ppm of output} + 0.1 \text{ ppm of range})$ , full load to no load

**Line Regulation** .....  $<0.1 \text{ ppm change, } \pm 10\%$  of selected nominal line

**Settling Time** ..... 3 seconds to full specification; + 1 second for range or polarity change; + 1 second for 1100 V range

**Overshoot** ..... <5 %

**Common Mode Rejection** ..... 140 dB, DC to 400 Hz

**Remote Sensing** ..... Available 0 V to  $\pm 1100$  V, on 2.2 V through 1100 V ranges

## AC Voltage Specifications

5730A AC Voltage Specifications: 99 % Confidence Level

Range	Resolution	Frequency (Hz)	Absolute / $\pm 5$ °C from calibration temperature				Relative $\pm 1$ °C	
			24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days
			$\pm$ (ppm output + $\mu$ V)					
2.2 mV	1 nV	10 - 20	250 + 5	270 + 5	290 + 5	300 + 5	250 + 5	270 + 5
		20 - 40	100 + 5	105 + 5	110 + 5	115 + 5	100 + 5	105 + 5
		40 - 20 k	85 + 5	90 + 5	95 + 5	100 + 5	60 + 5	65 + 5
		20 k - 50 k	220 + 5	230 + 5	240 + 5	250 + 5	85 + 5	95 + 5
		50 k - 100 k	500 + 6	540 + 6	570 + 6	600 + 6	200 + 6	220 + 6
		100 k - 300 k	1000 + 12	1200 + 12	1250 + 12	1300 + 12	350 + 12	400 + 12
		300 k - 500 k	1400 + 25	1500 + 25	1600 + 25	1700 + 25	800 + 25	1000 + 25
22 mV	10 nV	500 k - 1 M	2900 + 25	3100 + 25	3250 + 25	3400 + 25	2700 + 25	3000 + 25
		10 - 20	250 + 5	270 + 5	290 + 5	300 + 5	250 + 5	270 + 5
		20 - 40	100 + 5	105 + 5	110 + 5	115 + 5	100 + 5	105 + 5
		40 - 20 k	85 + 5	90 + 5	95 + 5	100 + 5	60 + 5	65 + 5
		20 k - 50 k	220 + 5	230 + 5	240 + 5	250 + 5	85 + 5	95 + 5
		50 k - 100 k	500 + 6	540 + 6	570 + 6	600 + 6	200 + 6	220 + 6
		100 k - 300 k	1000 + 12	1200 + 12	1250 + 12	1300 + 12	350 + 12	400 + 12
220 mV	100 nV	300 k - 500 k	1400 + 25	1500 + 25	1600 + 25	1700 + 25	800 + 25	1000 + 25
		500 k - 1 M	2900 + 25	3100 + 25	3250 + 25	3400 + 25	2700 + 25	3000 + 25
		10 - 20	250 + 15	270 + 15	290 + 15	300 + 15	250 + 15	270 + 15
		20 - 40	100 + 8	105 + 8	110 + 8	115 + 8	100 + 8	105 + 8
		40 - 20 k	65 + 8	66 + 8	67 + 8	70 + 8	60 + 8	65 + 8
		20 k - 50 k	135 + 8	140 + 8	145 + 8	150 + 8	85 + 8	95 + 8
		50 k - 100 k	370 + 20	380 + 20	390 + 20	400 + 20	200 + 20	220 + 20
2.2 V	1 $\mu$ V	100 k - 300 k	650 + 25	700 + 25	750 + 25	800 + 25	350 + 25	400 + 25
		300 k - 500 k	1400 + 30	1500 + 30	1600 + 30	1700 + 30	800 + 30	1000 + 30
		500 k - 1 M	2700 + 60	2900 + 60	3100 + 60	3300 + 60	2600 + 60	2800 + 60
		10 - 20	250 + 50	270 + 50	290 + 50	300 + 50	250 + 50	270 + 50
		20 - 40	95 + 20	100 + 20	105 + 20	110 + 20	95 + 20	100 + 20
		40 - 20 k	45 + 10	46 + 10	47 + 10	48 + 10	30 + 10	40 + 10
		20 k - 50 k	75 + 12	77 + 12	78 + 12	80 + 12	70 + 12	75 + 12

22 V	10 $\mu$ V	10 - 20	250 + 500	270 + 500	290 + 500	300 + 500	250 + 500	270 + 500	
		20 - 40	95 + 200	100 + 200	105 + 200	110 + 200	95 + 200	100 + 200	
		40 - 20 k	45 + 70	46 + 70	47 + 70	48 + 70	30 + 70	40 + 70	
		20 k - 50 k	75 + 120	77 + 120	78 + 120	80 + 120	70 + 120	75 + 120	
		50 k - 100 k	95 + 250	97 + 250	98 + 250	100 + 250	100 + 250	105 + 250	
		100 k - 300 k	285 + 800	290 + 800	295 + 800	300 + 800	270 + 800	290 + 800	
		300 k - 500 k	1000 + 2500	1100 + 2500	1150 + 2500	1200 + 2500	900 + 2500	1000 + 2500	
		500 k - 1 M	1500 + 4000	1600 + 4000	1700 + 4000	1800 + 4000	1300 + 4000	1400 + 4000	
		$\pm$ (ppm output + mV)							
220 V <sup>[2]</sup>	100 $\mu$ V	10 - 20	250 + 5	270 + 5	290 + 5	300 + 5	250 + 5	270 + 5	
		20 - 40	95 + 2	100 + 2	105 + 2	110 + 2	95 + 2	100 + 2	
		40 - 20 k	57 + 0.7	60 + 0.7	62 + 0.7	65 + 0.7	45 + 0.7	50 + 0.7	
		20 k - 50 k	90 + 1.2	95 + 1.2	97 + 1.2	100 + 1.2	75 + 1.2	80 + 1.2	
		50 k - 100 k	160 + 3	170 + 3	175 + 3	180 + 3	140 + 3	150 + 3	
		100 k - 300 k	900 + 20	1000 + 20	1050 + 20	1100 + 20	600 + 20	700 + 20	
		300 k - 500 k	5000 + 50	5200 + 50	5300 + 50	5400 + 50	4500 + 50	4700 + 50	
		500 k - 1 M	8000 + 100	9000 + 100	9500 + 100	10,000 + 100	8000 + 100	8500 + 100	
		$\pm$ (ppm output + mV)							
1100 V <sup>[1]</sup>	1 mV	15 - 50	300 + 20	320 + 20	340 + 20	360 + 20	300 + 20	320 + 20	
5725A Amplifier:									
1100 V	1 mV	40 - 1 k	75 + 4	80 + 4	85 + 4	90 + 4	50 + 4	55 + 4	
		1 k - 20 k	105 + 6	125 + 6	135 + 6	165 + 6	85 + 6	105 + 6	
		20 k - 30 k	230 + 11	360 + 11	440 + 11	600 + 11	160 + 11	320 + 11	
750 V		30 k - 50 k	230 + 11	360 + 11	440 + 11	600 + 11	160 + 11	320 + 11	
		50 k - 100 k	600 + 45	1300 + 45	1600 + 45	2300 + 45	380 + 45	1200 + 45	
Notes:									
1. Maximum output 250 V from 15-50 Hz.									
2. See Volt-Hertz capability in Figure A.									

**5730A AC Voltage Specifications: 95 % Confidence Level**

Range	Resolution	Frequency (Hz)	Absolute / $\pm$ 5 °C from calibration temperature				Relative $\pm$ 1 °C	
			24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days
			$\pm$ (ppm output + $\mu$ V)					
2.2 mV	1 nV	10 - 20	200 + 4	220 + 4	230 + 4	240 + 4	200 + 4	220 + 4
		20 - 40	80 + 4	85 + 4	87 + 4	90 + 4	80 + 4	85 + 4
		40 - 20 k	70 + 4	75 + 4	77 + 4	80 + 4	50 + 4	55 + 4
		20 k - 50 k	170 + 4	180 + 4	190 + 4	200 + 4	70 + 4	80 + 4
		50 k - 100 k	400 + 5	460 + 5	480 + 5	500 + 5	160 + 5	180 + 5
		100 k - 300 k	800 + 10	900 + 10	1000 + 10	1050 + 10	280 + 10	320 + 10
		300 k - 500 k	1100 + 20	1200 + 20	1300 + 20	1400 + 20	650 + 20	800 + 20
		500 k - 1 M	2400 + 20	2500 + 20	2600 + 20	2700 + 20	2100 + 20	2400 + 20
		$\pm$ (ppm output + $\mu$ V)						
22 mV	10 nV	10 - 20	200 + 4	220 + 4	230 + 4	240 + 4	200 + 4	220 + 4
		20 - 40	80 + 4	85 + 4	87 + 4	90 + 4	80 + 4	85 + 4
		40 - 20 k	70 + 4	75 + 4	77 + 4	80 + 4	50 + 4	55 + 4
		20 k - 50 k	170 + 4	180 + 4	190 + 4	200 + 4	70 + 4	80 + 4
		50 k - 100 k	400 + 5	460 + 5	480 + 5	500 + 5	160 + 5	180 + 5
		100 k - 300 k	800 + 10	900 + 10	1000 + 10	1050 + 10	280 + 10	320 + 10
		300 k - 500 k	1100 + 20	1200 + 20	1300 + 20	1400 + 20	650 + 20	800 + 20
		500 k - 1 M	2400 + 20	2500 + 20	2600 + 20	2700 + 20	2100 + 20	2400 + 20
		$\pm$ (ppm output + $\mu$ V)						
220 mV	100 nV	10 - 20	200 + 12	220 + 12	230 + 12	240 + 12	200 + 12	220 + 12
		20 - 40	80 + 7	85 + 7	87 + 7	90 + 7	80 + 7	85 + 7
		40 - 20 k	54 + 7	55 + 7	56 + 7	57 + 7	50 + 7	55 + 7
		20 k - 50 k	105 + 7	110 + 7	115 + 7	120 + 7	70 + 7	80 + 7
		50 k - 100 k	296 + 17	298 + 17	303 + 17	310 + 17	160 + 17	180 + 17
		100 k - 300 k	535 + 20	583 + 20	600 + 20	655 + 20	280 + 20	320 + 20
		300 k - 500 k	1100 + 25	1200 + 25	1300 + 25	1400 + 25	650 + 25	800 + 25
		500 k - 1 M	2400 + 45	2500 + 45	2600 + 45	2700 + 45	2100 + 45	2400 + 45
		$\pm$ (ppm output + $\mu$ V)						
2.2 V	1 $\mu$ V	10 - 20	200 + 40	220 + 40	230 + 40	240 + 40	200 + 40	220 + 40
		20 - 40	75 + 15	80 + 15	85 + 15	90 + 15	75 + 15	80 + 15
		40 - 20 k	37 + 8	39 + 8	40 + 8	42 + 8	25 + 8	35 + 8
		20 k - 50 k	61 + 10	63 + 10	65 + 10	67 + 10	55 + 10	60 + 10
		50 k - 100 k	79 + 30	81 + 30	82 + 30	85 + 30	80 + 30	85 + 30
		100 k - 300 k	276 + 80	300 + 80	314 + 80	336 + 80	230 + 80	250 + 80
		300 k - 500 k	800 + 200	900 + 200	950 + 200	1000 + 200	700 + 200	800 + 200
		500 k - 1 M	1300 + 300	1500 + 300	1600 + 300	1700 + 300	1000 + 300	1100 + 300
		$\pm$ (ppm output + $\mu$ V)						

22 V	10 $\mu$ V	10 - 20 20 - 40 40 - 20k 20k - 50k 50k - 100k 100k - 300k 300k - 500k 500k - 1M	200 + 400 75 + 150 37 + 50 61 + 100 78 + 200 238 + 600 800 + 2000 1200 + 3200	220 + 400 80 + 150 39 + 50 63 + 100 80 + 200 243 + 600 900 + 2000 1300 + 3200	230 + 400 85 + 150 40 + 50 65 + 100 81 + 200 249 + 600 900 + 2000 1400 + 3200	240 + 400 90 + 150 42 + 50 67 + 100 83 + 200 254 + 600 1000 + 2000 1500 + 3200	200 + 400 75 + 150 25 + 50 55 + 100 80 + 200 250 + 600 700 + 2000 1100 + 3200	220 + 400 80 + 150 35 + 50 60 + 100 85 + 200 270 + 600 800 + 2000 1200 + 3200
±(ppm output + mV)								
220 V <sup>[2]</sup>	100 $\mu$ V	10 - 20 20 - 40 40 - 20 k 20 k - 50 k 50 k - 100 k 100 k - 300 k 300 k - 500 k 500 k - 1 M	200 + 4 75 + 1.5 45 + 0.6 70 + 1 120 + 2.5 700 + 16 4000 + 40 6000 + 80	220 + 4 80 + 1.5 47 + 0.6 75 + 1 130 + 2.5 800 + 16 4200 + 40 7000 + 80	230 + 4 85 + 1.5 50 + 0.6 77 + 1 140 + 2.5 850 + 16 4300 + 40 7500 + 80	240 + 4 90 + 1.5 52 + 0.6 80 + 1 150 + 2.5 900 + 16 4400 + 40 8000 + 80	200 + 4 75 + 1.5 35 + 0.6 60 + 1 110 + 2.5 500 + 16 3600 + 40 6500 + 80	220 + 4 80 + 1.5 40 + 0.6 65 + 1 120 + 2.5 600 + 16 3800 + 40 7000 + 80
1100 V <sup>[1]</sup>	1 mV	15 - 50 50 - 1 k	240 + 16 55 + 3.5	260 + 16 60 + 3.5	280 + 16 65 + 3.5	300 + 16 70 + 3.5	240 + 16 40 + 3.5	260 + 16 45 + 3.5
<b>5725A Amplifier:</b>								
1100 V	1 mV	40 - 1 k 1 k - 20 k 20 k - 30 k	75 + 4 105 + 6 230 + 11	80 + 4 125 + 6 360 + 11	85 + 4 135 + 6 440 + 11	90 + 4 165 + 6 600 + 11	50 + 4 85 + 6 160 + 11	55 + 4 105 + 6 320 + 11
750 V		30 k - 50 k 50 k - 100 k	230 + 11 600 + 45	360 + 11 1300 + 45	440 + 11 1600 + 45	600 + 11 2300 + 45	160 + 11 380 + 45	320 + 11 1200 + 45
Notes: 1. Maximum output 250 V from 15-50 Hz. 2. See Volt-Hertz capability in Figure A.								

**AC Voltage Secondary Performance Specifications and Operating Characteristics**

Range	Frequency (Hz)	Stability $\pm 1^{\circ}\text{C}$ <sup>[1]</sup> 24 Hours	Temperature Coefficient		Output Impedance ( $\Omega$ )	Maximum Distortion Bandwidth 10 Hz-10 MHz	
			$\pm \mu\text{V}$	$\pm \mu\text{V} / ^{\circ}\text{C}$		$\pm (\text{ppm output} + \mu\text{V})$	$\pm (\text{ppm output} + \mu\text{V}) / ^{\circ}\text{C}$
2.2 mV	10 - 20	5	0.05	0.05	50	0.05 + 10	0.035 + 10
	20 - 40		0.05	0.05			
	40 - 20 k		0.05	0.05			
	20 k - 50 k		0.1	0.1			
	50 k - 100 k		0.2	0.2			
	100 k - 300 k		0.3	0.3			
	300 k - 500 k		0.4	0.4			
	500 k - 1 M		0.5	0.5			
22 mV	10 - 20	5	0.2	0.3	50	0.05 + 11	0.035 + 11
	20 - 40		0.2	0.3			
	40 - 20 k		0.2	0.3			
	20 k - 50 k		0.4	0.5			
	50 k - 100 k		0.5	0.5			
	100 k - 300 k		0.6	0.6			
	300 k - 500 k		1	1			
	500 k - 1 M		1	1			
220 mV	10 - 20	150 + 20 80 + 15 12 + 2 10 + 2 10 + 2 20 + 4 100 + 10 200 + 20	2 + 1	2 + 1	50	0.05 + 16	0.035 + 16
	20 - 40		2 + 1	2 + 1			
	40 - 20 k		2 + 1	2 + 1			
	20 k - 50 k		15 + 2	15 + 2			
	50 k - 100 k		15 + 4	15 + 4			
	100 k - 300 k		80 + 5	80 + 5			
	300 k - 500 k		80 + 5	80 + 5			
	500 k - 1 M		80 + 5	80 + 5			
2.2 V	10 - 20	150 + 20 80 + 15 12 + 4 15 + 5 15 + 5 30 + 10 70 + 20 150 + 50	50 + 10	50 + 10	10 + 2 10 + 2 5 + 2 15 + 4 20 + 4 80 + 15 80 + 40 80 + 100	0.05 + 80 0.035 + 80 0.035 + 80 0.035 + 80 120 + 16 300 ppm 600 ppm 1200 ppm	0.035 + 80 0.035 + 80 0.035 + 80 0.035 + 80 0.035 + 110 0.3 + 110 0.5 + 110 1 + 110
	20 - 40		15 + 5	15 + 5			
	40 - 20 k		2 + 1	5 + 2			
	20 k - 50 k		10 + 2	15 + 4			
	50 k - 100 k		10 + 4	20 + 4			
	100 k - 300 k		80 + 15	80 + 15			
	300 k - 500 k		80 + 40	80 + 40			
	500 k - 1 M		80 + 100	80 + 100			

	10 - 20	150 + 20	50 + 100	50 + 100	10 + 20	0.05 + 700
	20 - 40	80 + 15	15 + 30	15 + 40	10 + 20	0.035 + 700
	40 - 20 k	12 + 8	2 + 10	4 + 15	10 + 30	0.035 + 700
	20 k - 50 k	15 + 10	10 + 20	20 + 20	30 + 50	0.035 + 700
	50 k - 100 k	15 + 10	10 + 40	20 + 40	80 + 80	0.05 + 800
	100 k - 300 k	30 + 15	80 + 150	80 + 150	100 + 700	0.3 + 800
	300 k - 500 k	70 + 100	80 + 300	80 + 300	200 + 1100	0.3 + 800
	500 k - 1 M	150 + 100	80 + 500	80 + 500	600 + 3000	2 + 800
22 V						
	10 - 20	150 + 200	50 + 1000	50 + 1000	10 + 200	0.05 + 10,000
	20 - 40	80 + 150	15 + 300	15 + 300	10 + 200	0.05 + 10,000
	40 - 20 k	12 + 80	2 + 80	4 + 80	10 + 300	0.05 + 10,000
	20 k - 50 k	15 + 100	10 + 100	20 + 100	30 + .600	0.05 + 10,000
	50 k - 100 k	15 + 100	10 + 500	20 + 500	80 + 3,000	0.2 + 50,000
	100 k - 300 k	30 + 400	80 + 600	80 + 600	250 + 25,000	1.5 + 50,000
	300 k - 500 k	100 + 10,000	80 + 800	80 + 800	500 + 50,000	1.5 + 50,000
	500 k - 1 M	200 + 20,000	80 + 1000	80 + 1000	1000 + 110,000	3.5 + 100,000
		±(ppm output + mV)	±(ppm output) / °C	±(ppm output + mV)	±(% output)	
1100 V	15 - 50	150 + 0.5	50	50	10 + 2	0.15
	50 - 1 k	20 + 0.5	2	5	10 + 1	0.07




Figure A.

5725A Amplifier:							
Range	Frequency (Hz)	Stability $\pm 1^\circ\text{C}$ <sup>[1]</sup> 24 Hours	Temperature Coefficient Adder		Load Regulation <sup>[2]</sup>	Distortion Bandwidth 10 Hz -10 MHz $\pm \{ \%$ output)	
			10 - 40 °C	0 - 10 °C and 40 - 50 °C		150 pF	1000 pF
		$\pm (\text{ppm output} + \text{mV})$	$\pm (\text{ppm output}) / ^\circ\text{C}$			$\pm (\text{ppm output} + \text{mV})$	$\pm (\text{ppm output} + \text{mV})$
1100 V	40 - 1 k	10 + .5	5	5	10 + 1	0.10	0.10
	1 k - 20 k	15 + 2	5	5	90 + 6	0.10	0.15
	20 k - 50 k	40 + 2	10	10	275 + 11	0.30	0.30
	50 k - 100 k	130 + 2	30	30	500 + 30	0.40	0.40

## Notes:

1. Stability specifications are included in Absolute specification values for the primary specifications.
2. The 5725A will drive up to 1000 pF of load capacitance. Absolute specifications include loads to 300 pF and 150 pF as shown under "Load Limits." For capacitances up to the maximum of 1000 pF, add "Load Regulation."

Voltage Range	Maximum Current Limits		Load Limits
2.2 V <sup>[2]</sup>	50 mA, 0 °C-40 °C 20 mA, 40 °C-50 °C		>50 Ω, 1000 pF
22 V			
220 V			
1100 V	6 mA		600 pF
5725A Amplifier:			
1100 V	40 Hz-5 kHz	50 mA	1000 pF <sup>[1]</sup>
	5 kHz-30 kHz	70 mA	300 pF
	30 kHz-100 kHz	70 mA <sup>[3]</sup>	150 pF

## Notes:

1. The 5725A will drive up to 1000 pF of load capacitance. Absolute specifications include loads to 300 pF and 150 pF as shown under "Load Limits." For capacitances up to the maximum of 1000 pF, add "Load Regulation."
2. 2.2 V Range, 100 kHz-1.2 MHz only: Absolute specifications cover loads to 10 mA or 1000 pF. For higher loads, load regulation is added.
3. Applies from 0 °C to 40 °C.

**Output Display Formats** ..... Voltage or dBm, dBm reference 600 Ω.

**Minimum Output** ..... 10 % on each range

**External Sense** ..... Applicable for 2.2 V, 22 V, 220 V, and 1100 V ranges; 5730A  
< 100 kHz, 5725A < 30 kHz. Specifications are the same as internal sense.

#### Settling Time to Published Specifications

Frequency (Hz)	Settling Time (seconds)
10-120	7
>120	5

## Notes:

Plus 1 second for amplitude or frequency range change

Plus 2 seconds for 5730A 1100 V range

Plus 4 seconds for 5725A 1100 V range

**Overshoot** ..... < 10 %

**Common Mode Rejection** ..... 140 dB, dc to 400 Hz

#### Frequency

Ranges (Hz) ..... 10.000 - 119.99  
0.1200 k - 1.1999 k  
1.200 k - 11.999 k  
12.00 k - 119.99 k  
120.0 k - 1.1999 M

Absolute Specification ..... ±0.0025 %

Resolution ..... 11.999 counts

**Phase Lock (Selectable Rear Panel BNC Input)**Phase Specification (except 1100 V range) .... >30 Hz:  $\pm 1^\circ + 0.05^\circ/\text{kHz}$ , <30 Hz:  $\pm 3^\circ$ 

Input Voltage ..... 1 V to 10 V rms sine wave (do not exceed 1 V for mV ranges)

Frequency Range ..... 10 Hz to 1.1999 MHz

Lock Range .....  $\pm 2\%$  of frequency

Lock-In Time ..... Larger of 10/frequency or 10 msec

**Phase Reference (Selectable Rear Panel BNC Output)**Range .....  $\pm 180^\circ$ Phase Absolute Specification ..... (except 1100 V range)  $\pm 1^\circ$  at quadrature points ( $0^\circ, \pm 90^\circ, \pm 180^\circ$ )  
elsewhere  $\pm 2^\circ$ Stability .....  $\pm 0.1^\circ$ 

Resolution ..... 1°

Output Level ..... 2.5 V rms  $\pm 0.2$  V

Frequency Range ..... 50 Hz to 1 kHz, usable 10 Hz to 1.1999 MHz

**Resistance Specifications****5730A Resistance Specifications**

Nominal Value ( $\Omega$ )	Absolute Specification of Characterized Value $\pm 5^\circ \text{C}$ from calibration temperature <sup>[1]</sup>					Relative $\pm 1^\circ \text{C}$	
	24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days	
	$\pm \text{ppm}$						
99 % Confidence Level							
0	50 $\mu\Omega$	50 $\mu\Omega$	50 $\mu\Omega$	50 $\mu\Omega$	50 $\mu\Omega$	50 $\mu\Omega$	50 $\mu\Omega$
1	85	95	100	110	32	40	
1.9	85	95	100	110	25	33	
10	23	25	26	27	5	8	
19	23	25	26	27	4	7	
100	10	11	11.5	12	2	4	
190	10	11	11.5	12	2	4	
1 k	7	7.2	7.5	8	2	3	
1.9 k	7	7.2	7.5	8	2	3	
10 k	6	7	7.5	8	2	3	
19 k	6	7	7.5	8	2	3	
100 k	7	8	9	10	2	3	
190 k	8	10	11	12	2	3	
1 M	13	14	14.5	15	2.5	5	
1.9 M	15	17	19	21	3	6	
10 M	33	37	40	46	10	14	
19 M	43	47	50	55	20	24	
100 M	100	110	115	120	50	60	
95 % Confidence Level							
0	40 $\mu\Omega$	40 $\mu\Omega$	40 $\mu\Omega$	40 $\mu\Omega$	40 $\mu\Omega$	40 $\mu\Omega$	40 $\mu\Omega$
1	70	80	85	95	27	35	
1.9	70	80	85	95	20	26	
10	20	21	22	23	4	7	
19	20	21	22	23	3.5	6	
100	8	9	9.5	10	1.6	3.5	
190	8	9	9.5	10	1.6	3.5	
1 k	5.5	5.7	6	6.5	1.6	2.5	
1.9 k	5.5	5.7	6	6.5	1.6	2.5	
10 k	5	5.5	6	6.5	1.6	2.5	
19 k	5	5.5	6	6.5	1.6	2.5	
100 k	5.5	7.5	8	8.5	1.6	2.5	
190 k	6	7	8	8.5	1.6	2.5	
1 M	10	11	12	13	2	4	
1.9 M	12	13.5	15	18	2.5	4	
10 M	27	31	34	40	8	12	
19 M	35	39	42	47	16	20	
100 M	85	95	100	100	40	50	

Note:

1. Specifications apply to displayed value. 4-wire connections, except 100 M $\Omega$ .

## Resistance Secondary Performance Specifications and Operating Characteristics

Nominal Value (Ω)	Stability ±1 °C [1] 24 Hours	Temperature Coefficient Adder [2]		Full Spec Load Range [3] $I_L - I_U$ (mA)	Maximum Peak Current $I_{MAX}$ (mA)	Maximum Difference of Characterized to Nominal Value	Two-Wire Adder Active Compensation [4]	
		10 - 40 °C	0 - 10 °C and 40 - 50 °C				Lead Resistance	
		±ppm	±ppm/°C				±ppm	0.1 Ω 1 Ω
0	—	—	—	8 - 500	500	—	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
1	32	4	5	8 - 100	700	500	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
1.9	25	6	7	8 - 100	500	500	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
10	5	2	3	8 - 11	220	300	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
19	4	2	3	8 - 11	160	300	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
100	2	2	3	8 - 11	70	150	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
190	2	2	3	8 - 11	50	150	$2 + \frac{4\mu V}{I_m}$	$4 + \frac{4\mu V}{I_m}$
1 k	2	2	3	1 - 2	22	150	10	15
1.9 k	2	2	3	1 - 1.5	16	150	10	15
10 k	2	2	3	100 - 500 μA	7	150	50	60
19 k	2	2	3	50 - 250 μA	5	150	100	120
100 k	2	2	3	10 - 100 μA	1	150	$I_m = \text{Current produced by Ohmmeter (A)}$	
190 k	2	2	3	5 - 100 μA	500 μA	150		
1 M	2.5	2.5	6	5 - 20 μA	100 μA	200		
1.9 M	3.5	3	10	2.5 - 10 μA	50 μA	200		
10 M	10	5	20	0.5 - 2 μA	10 μA	300		
19 M	20	8	40	0.25 - 1 μA	5 μA	300		
100 M	50	12	100	50 - 200 nA	1 μA	500		

## Notes:

1. Stability specifications are included in the Absolute specification values in the primary specification tables.
2. Temperature coefficient is an adder to Absolute specifications that does not apply unless operated more than 5 °C from calibration temperature, or calibrated outside the range 19 °C to 24 °C. Two examples:
  - Calibrate at 20 °C: Temperature coefficient adder is not required unless operated below 15 °C or above 25 °C.
  - Calibrate at 26 °C: Add 2 °C temperature coefficient adder. Additional temperature coefficient adder is not required unless operated below 21 °C or above 31 °C.
3. Refer to current derating factors table for loads outside of this range.
4. Active two-wire compensation may be selected for values less than 100 kΩ, with either the front panel or the meter input terminals as reference plane. Active compensation is limited to 11 mA load, and to 2 V burden. Two-wire compensation can be used only with Ω-meters that source continuous (not pulsed) dc current.

**Current Derating Factors**

<b>Nominal Value (Ω)</b>	<b>Value of Derating Factor K for Over or Under Current</b>		
	<b>Two-Wire Comp <math>I &lt; I_L^{[1]}</math></b>	<b>Four-Wire <math>I &lt; I_L^{[1]}</math></b>	<b>Four-Wire <math>I_U &lt; I &lt; I_{MAX}^{[2]}</math></b>
SHORT	<b>4.4</b>	0.3	—
1	<b>4.4</b>	300	$4 \times 10^{-5}$
1.9	<b>4.4</b>	160	$1.5 \times 10^{-4}$
10	<b>4.4</b>	30	$1.6 \times 10^{-3}$
19	<b>4.4</b>	16	$3 \times 10^{-3}$
100	<b>4.4</b>	3.5	$1 \times 10^{-2}$
190	<b>4.4</b>	2.5	$1.9 \times 10^{-2}$
1 k	<b>4.4</b>	0.4	0.1
1.9 k	<b>4.4</b>	0.4	0.19
10 k	5000	50	2.0
19 k	5000	50	3.8
100 k	—	7.5	$2 \times 10^{-5}$
190 k	—	4.0	$3.8 \times 10^{-5}$
1 M	—	1.0	$1.5 \times 10^{-4}$
1.9 M	—	0.53	$2.9 \times 10^{-4}$
10 M	—	0.2	$1 \times 10^{-3}$
19 M	—	0.53	$1.9 \times 10^{-3}$
100 M	—	0.1	—

## Notes:

1. For  $I < I_L$ , errors occur due to thermally generated voltages within the 5730A. Use the following equation to determine the error, and add this error to the corresponding specifications.

$$\text{Error} = K(I_L - I)/(I_L \times I)$$

Where: Error is in mΩ for all two-wire comp values and four-wire short, and in ppm for the remaining four-wire values.

K is the constant from the above table;

I and  $I_L$  are expressed in mA for short to 1.9 kΩ;

I and  $I_L$  are expressed in μA for 10 kΩ to 100 MΩ

2. For  $I_U < I < I_{MAX}$  errors occur due to self-heating of the resistors in the calibrator. Use the following equation to determine the error in ppm and add this error to the corresponding specifications.

$$\text{Error in ppm} = K(I^2 - I_U^2)$$

Where: K is the constant from the above table;

I and  $I_U$  are expressed in mA for short to 19 kΩ;

I and  $I_U$  are expressed in μA for 100 kΩ to 100 MΩ

## DC Current Specifications

### 5730A DC Current Specifications

Range	Resolution	Absolute / $\pm 5^\circ\text{C}$ from calibration temperature				Relative $\pm 1^\circ\text{C}$								
		24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days							
		nA	$\pm(\text{ppm output} + \text{nA})$											
<b>99 % Confidence Level</b>														
220 $\mu\text{A}$	0.1	40 + 7	42 + 7	45 + 7	50 + 7	24 + 2	26 + 2							
2.2 mA	1	30 + 8	35 + 8	37 + 8	40 + 8	24 + 5	26 + 5							
22 mA	10	30 + 50	35 + 50	37 + 50	40 + 50	24 + 50	26 + 50							
	$\mu\text{A}$	$\pm(\text{ppm output} + \mu\text{A})$												
220 mA <sup>[1]</sup>	0.1	40 + 0.8	45 + 0.8	47 + 0.8	50 + 0.8	26 + 0.3	30 + 0.3							
2.2 A <sup>[1]</sup>	1	60 + 15	70 + 15	80 + 15	90 + 15	40 + 7	45 + 7							
<b>5725A Amplifier:</b>														
11 A	10	330 + 470	340 + 480	350 + 480	360 + 480	100 + 130	110 + 130							
<b>95 % Confidence Level</b>														
	nA	$\pm(\text{ppm output} + \text{nA})$												
220 $\mu\text{A}$	0.1	32 + 6	35 + 6	37 + 6	40 + 6	20 + 1.6	22 + 1.6							
2.2 mA	1	25 + 7	30 + 7	33 + 7	35 + 7	20 + 4	22 + 4							
22 mA	10	25 + 40	30 + 40	33 + 40	35 + 40	20 + 40	22 + 40							
	$\mu\text{A}$	$\pm(\text{ppm output} + \mu\text{A})$												
220 mA <sup>[1]</sup>	0.1	35 + 0.7	40 + 0.7	42 + 0.7	45 + 0.7	22 + 0.25	25 + 0.25							
2.2 A <sup>[1]</sup>	1	50 + 12	60 + 12	70 + 12	80 + 12	32 + 6	40 + 6							
<b>5725A Amplifier:</b>														
11 A	10	330 + 470	340 + 480	350 + 480	360 + 480	100 + 130	110 + 130							
Note:														
Maximum output from the calibrator's terminals is 2.2 A. Specifications for 220 $\mu\text{A}$ and 2.2 mA ranges are increased by a factor of 1.3 when supplied through 5725A terminals.														
Specifications are otherwise identical for all output locations.														
1.	Add to specifications:													
$\pm 200 \times 1^2 \text{ ppm}$ for >100 mA on 220 mA range														
$\pm 10 \times 1^2 \text{ ppm}$ for >1 A on 2.2 A range														

### DC Current Secondary Performance Specifications and Operating Characteristics

Range	Stability $\pm 1^\circ\text{C}$ <sup>[1]</sup> 24 Hours	Temperature Coefficient <sup>[2]</sup>		Compliance Limits	Burden Voltage Adder <sup>[3]</sup> ( $\pm \text{nA/V}$ )	Maximum Load for Published Specification <sup>[4]</sup> ( $\Omega$ )	Noise	
		10 - 40 °C	0 - 10 °C and 40 - 50 °C				Bandwidth 0.1-10 Hz	Bandwidth 10 Hz-10 kHz
		$\pm(\text{ppm}$ $\text{output} + \text{nA})$	$\pm(\text{ppm output} + \text{nA}) /$ $^\circ\text{C}$				pk-pk	RMS
220 $\mu\text{A}$	5 + 1	1 + 0.40	3 + 1	10	0.2	20k	6 + .9	10
2.2 mA	5 + 5	1 + 2	3 + 10	10	0.2	2k	6 + 5	10
22 mA	5 + 50	1 + 20	3 + 100	10	10	200	6 + 50	50
220 mA	8 + 300	1 + 200	3 + 1 $\mu\text{A}$	10	100	20	9 + 300	500
2.2 A	9 + 7 $\mu\text{A}$	1 + 2.5 $\mu\text{A}$	3 + 10 $\mu\text{A}$	3 <sup>[5]</sup>	2 $\mu\text{A}$	2	12 + 1.5 $\mu\text{A}$	20 $\mu\text{A}$
<b>5725A</b>	$\pm(\text{ppm}$ $\text{output} + \mu\text{A})$	$\pm(\text{ppm output} + \mu\text{A}) /$ $^\circ\text{C}$					$\pm(\text{ppm output} + \mu\text{A})$	$\mu\text{A}$
11 A	25 + 100	20 + 75	30 + 120	4	0	4	15 + 70	175
Notes:								
Maximum output from the calibrator's terminals is 2.2 A. Specifications for 220 $\mu\text{A}$ and 2.2 mA ranges are increased by a factor of 1.3 when supplied through 5725A terminals.								
1.	Stability specifications are included in the Absolute specification values for the primary specifications.							
2.	Temperature coefficient is an adder to Absolute specifications. It does not apply unless operating more than $\pm 5^\circ\text{C}$ from calibration temperature.							
3.	Burden voltage adder is an adder to Absolute specifications that does not apply unless burden voltage is greater than 0.5 V.							
4.	For higher loads, multiply Absolute specification by: $1 + \frac{\text{0.1 x actual load}}{\text{maximum load for published specification}}$							
5.	The calibrator's compliance limit is 2 V for outputs from 1 A to 2.2 A. 5725A Amplifier may be used in range-lock mode down to 0 A.							

- Minimum Output** ..... 0 for all ranges, including 5725A.
- Settling Time** ..... 1 second for  $\mu$ A and mA ranges; 3 seconds for 2.2 A range; 6 seconds for 11 range; + 1 second for range or polarity change
- Overshoot** ..... <5 %

## AC Current Specifications

5730A AC Current Specifications: 99 % Confidence Level

Range	Resolution	Frequency (Hz)	Absolute / $\pm 5^\circ\text{C}$ from calibration temperature				Relative $\pm 1^\circ\text{C}$	
			24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days
			$\pm(\text{ppm output} + \text{nA})$					
220 $\mu$ A	1 nA	10 - 20	260 + 20	280 + 20	290 + 20	300 + 20	260 + 20	280 + 20
		20 - 40	170 + 12	180 + 12	190 + 12	200 + 12	130 + 12	150 + 12
		40 - 1 k	115 + 10	117 + 10	118 + 10	120 + 10	100 + 10	110 + 10
		1 k - 5 k	300 + 15	320 + 15	340 + 15	350 + 15	250 + 15	280 + 15
		5 k - 10 k	1000 + 80	1100 + 80	1200 + 80	1300 + 80	900 + 80	1000 + 80
2.2 mA	10 nA	10 - 20	260 + 50	280 + 50	290 + 50	300 + 50	260 + 50	280 + 50
		20 - 40	170 + 40	180 + 40	190 + 40	200 + 40	130 + 40	150 + 40
		40 - 1 k	115 + 40	117 + 40	118 + 40	120 + 40	100 + 40	110 + 40
		1 k - 5 k	210 + 130	220 + 130	230 + 130	240 + 130	190 + 130	220 + 130
		5 k - 10 k	1000 + 800	1100 + 800	1200 + 800	1300 + 800	900 + 800	1000 + 800
22 mA	100 nA	10 - 20	260 + 500	280 + 500	290 + 500	300 + 500	260 + 500	280 + 500
		20 - 40	170 + 400	180 + 400	190 + 400	200 + 400	130 + 400	150 + 400
		40 - 1 k	115 + 400	117 + 400	118 + 400	120 + 400	100 + 400	110 + 400
		1 k - 5 k	210 + 700	220 + 700	230 + 700	240 + 700	190 + 700	220 + 700
		5 k - 10 k	1000 + 6000	1100 + 6000	1200 + 6000	1300 + 6000	900 + 6000	1000 + 6000
$\pm(\text{ppm output} + \mu\text{A})$								
220 mA	1 $\mu$ A	10 - 20	260 + 5	280 + 5	290 + 5	300 + 5	260 + 5	280 + 5
		20 - 40	170 + 4	180 + 4	190 + 4	200 + 4	130 + 4	150 + 4
		40 - 1 k	115 + 3	117 + 3	118 + 3	120 + 3	100 + 3	110 + 3
		1 k - 5 k	210 + 4	220 + 4	230 + 4	240 + 4	190 + 4	220 + 4
		5 k - 10 k	1000 + 12	1100 + 12	1200 + 12	1300 + 12	900 + 12	1000 + 12
2.2 A	10 $\mu$ A	20 - 1 k	270 + 40	280 + 40	290 + 40	300 + 40	260 + 40	280 + 40
		1 k - 5 k	440 + 100	460 + 100	480 + 100	500 + 100	420 + 100	440 + 100
		5 k - 10 k	6000 + 200	7000 + 200	7500 + 200	8000 + 200	6000 + 200	7000 + 200
<b>5725A Amplifier:</b>								
11 A	100 $\mu$ A	40 - 1 k	370 + 170	400 + 170	440 + 170	460 + 170	300 + 170	330 + 170
		1 k - 5 k	800 + 380	850 + 380	900 + 380	950 + 380	700 + 380	800 + 380
		5 k - 10 k	3000 + 750	3300 + 750	3500 + 750	3600 + 750	2800 + 750	3200 + 750

Note:

Maximum output from the calibrator's terminals is 2.2 A. Specifications for 220  $\mu$ A and 2.2 mA ranges are increased by a factor of 1.3 plus 2  $\mu$ A when supplied through 5725A terminals.

**5730A AC Current Specifications: 95 % Confidence Level**

Range	Resolution	Frequency (Hz)	Absolute / ±5 °C from calibration temperature				Relative ±1 °C	
			24 Hours	90 Days	180 Days	1 Year	24 Hours	90 Days
			±(ppm output + nA)					
220 μA	1 nA	10 - 20	210 + 16	230 + 16	240 + 16	250 + 16	210 + 16	230 + 16
		20 - 40	130 + 10	140 + 10	150 + 10	160 + 10	110 + 10	130 + 10
		40 - 1 k	96 + 8	99 + 8	101 + 8	103 + 8	80 + 8	90 + 8
		1 k - 5 k	240 + 12	250 + 12	270 + 12	280 + 12	200 + 12	230 + 12
		5 k - 10 k	800 + 65	900 + 65	1000 + 65	1100 + 65	700 + 65	800 + 65
2.2 mA	10 nA	10 - 20	210 + 40	230 + 40	240 + 40	250 + 40	210 + 40	230 + 40
		20 - 40	130 + 35	140 + 35	150 + 35	160 + 35	110 + 35	130 + 35
		40 - 1 k	96 + 35	99 + 35	101 + 35	103 + 35	80 + 35	90 + 35
		1 k - 5 k	170 + 110	180 + 110	190 + 110	200 + 110	160 + 110	170 + 110
		5 k - 10 k	800 + 650	900 + 650	1000 + 650	1100 + 650	700 + 650	800 + 650
22 mA	100 nA	10 - 20	210 + 400	230 + 400	240 + 400	250 + 400	210 + 400	230 + 400
		20 - 40	130 + 350	140 + 350	150 + 350	160 + 350	110 + 350	130 + 350
		40 - 1 k	96 + 350	99 + 350	101 + 350	103 + 350	80 + 350	90 + 350
		1 k - 5 k	170 + 550	180 + 550	190 + 550	200 + 550	160 + 550	170 + 550
		5 k - 10 k	800 + 5000	900 + 5000	1000 + 5000	1100 + 5000	700 + 5000	800 + 5000
			±(ppm output + μA)					
220 mA	1 μA	10 - 20	210 + 4	230 + 4	240 + 4	250 + 4	210 + 4	230 + 4
		20 - 40	130 + 3.5	140 + 3.5	150 + 3.5	160 + 3.5	110 + 3.5	130 + 3.5
		40 - 1 k	96 + 2.5	99 + 2.5	101 + 2.5	103 + 2.5	80 + 2.5	90 + 2.5
		1 k - 5 k	170 + 3.5	180 + 3.5	190 + 3.5	200 + 3.5	160 + 3.5	170 + 3.5
		5 k - 10 k	800 + 10	900 + 10	1000 + 10	1100 + 10	700 + 10	800 + 10
2.2 A	10 μA	20 - 1 k	214 + 35	224 + 35	234 + 35	244 + 35	200 + 35	230 + 35
		1 k - 5 k	350 + 80	390 + 80	420 + 80	450 + 80	300 + 80	350 + 80
		5 k - 10 k	5000 + 160	6000 + 160	6500 + 160	7000 + 160	5000 + 160	6000 + 160
<b>5725A Amplifier:</b>								
11 A	100 μA	40 - 1 k	370 + 170	400 + 170	440 + 170	460 + 170	300 + 170	330 + 170
		1 k - 5 k	800 + 380	850 + 380	900 + 380	950 + 380	700 + 380	800 + 38
		5 k - 10 k	3000 + 750	3300 + 750	3500 + 750	3600 + 750	2800 + 750	3200 + 750
<p>Note: Maximum output from the calibrator's terminals is 2.2 A. Specifications for 220 μA and 2.2 mA ranges are increased by 1.3 plus 2 μA when supplied through 5725A terminals.</p>								

## AC Current Secondary Performance Specifications and Operating Characteristics

Range	Frequency (Hz)	Stability $\pm 1^\circ\text{C}$ <sup>[1]</sup> 24 Hours	Temperature Coefficient <sup>[2]</sup>		Compliance Limits (V rms)	Maximum Resistive Load For Published Specification <sup>[3]</sup> ( $\Omega$ )	Noise and Distortion (Bandwidth 10 Hz - 50 kHz <0.5V Burden)
			10 - 40 °C	0 - 10 °C and 40 - 50 °C			
			$\pm(\text{ppm output} + \text{nA})$	$\pm(\text{ppm output} + \text{nA})/\text{°C}$			
220 $\mu\text{A}$	10 - 20	150 + 5	50 + 5	50 + 5	7	2 k	0.05 + 0.1
	20 - 40	80 + 5	20 + 5	20 + 5			0.05 + 0.1
	40 - 1 k	30 + 3	4 + 0.5	10 + 0.5			0.05 + 0.1
	1 k - 5 k	50 + 20	10 + 1	20 + 1			0.25 + 0.5
	5 k - 10 k	400 + 100	20 + 100	20 + 100			0.05 + 1
2.2 mA	10 - 20	150 + 5	50 + 5	50 + 5	7	800	0.05 + 0.1
	20 - 40	80 + 5	20 + 4	20 + 4			0.05 + 0.1
	40 - 1 k	30 + 3	4 + 1	10 + 2			0.05 + 0.1
	1 k - 5 k	50 + 20	10 + 100	20 + 100			0.25 + 0.5
	5 k - 10 k	400 + 100	50 + 400	50 + 400			0.05 + 1
22 mA	10 - 20	150 + 50	50 + 10	50 + 10	7	80	0.05 + 0.1
	20 - 40	80 + 50	20 + 10	20 + 10			0.05 + 0.1
	40 - 1 k	30 + 30	4 + 10	10 + 20			0.05 + 0.1
	1 k - 5 k	50 + 500	10 + 500	20 + 400			0.25 + 0.5
	5 k - 10 k	400 + 1000	50 + 1000	50 + 1000			0.05 + 1
	Hz	$\pm(\text{ppm output} + \mu\text{A})$	$\pm(\text{ppm output} + \mu\text{A})/\text{°C}$				
220 mA	10 - 20	150 + 0.5	50 + 0.05	50 + 0.05	7	8	0.05 + 10
	20 - 40	80 + 0.5	20 + 0.05	20 + 0.05			0.05 + 10
	40 - 1 k	30 + 0.3	4 + 0.1	10 + 0.1			0.05 + 10
	1 k - 5 k	50 + 3	10 + 2	20 + 2			0.25 + 50
	5 k - 10 k	400 + 5	50 + 5	50 + 5			0.05 + 100
2.2 A	20 - 1 k	50 + 5	4 + 1	10 + 1	1.4 <sup>[4]</sup>	0.8	0.5 + 100
	1 k - 5 k	80 + 20	10 + 5	20 + 5			0.3 + 500
	5 k - 10 k	800 + 50	50 + 10	50 + 10			1 + 1 mA
5725A Amplifier:							$\pm(\% \text{ output})$
11 A	40 - 1 k	75 + 100	20 + 75	30 + 75	3	3	0.05 <sup>[5]</sup>
	1 k - 5 k	100 + 150	40 + 75	50 + 75			0.12 <sup>[5]</sup>
	5 k - 10 k	200 + 300	100 + 75	100 + 75			0.5 <sup>[5]</sup>

## Notes:

Maximum output from 5730A terminals is 2.2 A. Specifications for 220  $\mu\text{A}$  and 2.2 mA ranges are increased by a factor of 1.3, plus 2  $\mu\text{A}$  when supplied through 5725A terminals. Specifications are otherwise identical for all output locations.

1. Stability specifications are included in the Absolute values for the primary specifications.
2. Temperature coefficient is an adder to specifications that does not apply unless operating more than  $\pm 5^\circ\text{C}$  from calibration temperature.
3. For larger resistive loads multiply accuracy specifications by:  $(\frac{\text{actual load}}{\text{maximum load for published specification}})^2$
4. 1.5 V compliance limit above 1 A. 5725A Amplifier may be used in range-lock mode down to 1 A.
5. For resistive loads within rated compliance voltage limits.

**Minimum Output**.....9  $\mu\text{A}$  for 220  $\mu\text{A}$  range, 10 % on all other ranges. 1 A minimum for 5725A.

**Inductive Load Limits**.....400  $\mu\text{H}$  (5730A, or 5725A). 20  $\mu\text{H}$  for 5730A output > 1 A.

**Power Factors**.....5730A, 0.9 to 1; 5725A, 0.1 to 1. Subject to compliance voltage limits.

**Frequency**

**Range (Hz)**.....10.000 - 11.999, 12.00 - 119.99, 120.0 - 1199.9, 1.200 k - 10.000 k

Specification ..... $\pm 0.01\%$  of output

Resolution .....11,999 counts

**Settling Time**.....5 seconds for 5730A ranges; 6 seconds for 5725A 11 A range;  
+1 second for amplitude or frequency range change.

**Overshoot**.....<10 %

## Wideband AC Voltage (Option 5730A/03 and 5730A/05) Specifications (99 % Confidence Level)

Specifications apply to the end of the cable and 50 Ω termination used for calibration.

Range		Resolution	Absolute / ±5 °C from calibration temperature 30 Hz - 500 kHz				
Volts	dBm		24 Hours	90 Days	180 Days	1 Year	
			±(% output + μV)				
1.1 mV	-46	10 nV	0.4 + 0.4	0.5 + 0.4	0.6 + 0.4	0.8 + 2	
3.3 mV	-37	10 nV	0.4 + 1	0.45 + 1	0.5 + 1	0.7 + 3	
11 mV	-26	100 nV	0.2 + 4	0.35 + 4	0.5 + 4	0.7 + 8	
33 mV	-17	100 nV	0.2 + 10	0.3 + 10	0.45 + 10	0.6 + 16	
110 mV	-6.2	1 μV	0.2 + 40	0.3 + 40	0.45 + 40	0.6 + 40	
330 mV	+3.4	1 μV	0.2 + 100	0.25 + 100	0.35 + 100	0.5 + 100	
1.1 V	+14	10 μV	0.2 + 400	0.25 + 400	0.35 + 400	0.5 + 400	
3.5 V	+24	10 μV	0.15 + 500	0.2 + 500	0.3 + 500	0.4 + 500	

Frequency (Hz)	Frequency Resolution (Hz)	Amplitude Flatness, 1 kHz Reference Voltage Range			Temperature Coefficient ±ppm/°C	Settling Time To Published Specification (Seconds)	Harmonic Distortion (dB)
		1.1 mV	3.3 mV	>3.3 mV			
		±(% output + floor indicated)					
10 - 30	0.01	0.3	0.3	0.3	100	7	-40
30 - 119.99	0.01	0.1	0.1	0.1	100	7	-40
120 - 1.1999 k	0.1	0.1	0.1	0.1	100	5	-40
1.2 k - 11.999 k	1	0.1	0.1	0.1	100	5	-40
12 k - 119.99 k	10	0.1	0.1	0.1	100	5	-40
120 k - 1.1999 M	100	0.2 + 3 μV	0.1 + 3 μV	0.1 + 3 μV	100	5	-40
1.2 M - 2 M <sup>[1]</sup>	1 k	0.2 + 3 μV	0.1 + 3 μV	0.1 + 3 μV	100	0.5	-40
2 M - 11.9 M	1 k	0.4 + 3 μV	0.3 + 3 μV	0.2 + 3 μV	100	0.5	-40
12 M - 20 M	10 k	0.6 + 3 μV	0.5 + 3 μV	0.4 + 3 μV	150	0.5	-34
20 M - 30 M	10 k	1.5 + 15 μV	1.5 + 3 μV	1 + 3 μV	300	0.5	-34
30 M - 50 M <sup>[2]</sup>	10 k	3.0 + 15 μV	3.0 + 3 μV	2.0 + 3 μV	600	0.5	-34

Note:

- For output voltages <50 % of full range in the 33 mV, 110 mV, 330 mV, 1.1 V, and 3.5 V ranges, add 0.1 % to the amplitude flatness specification.

### Additional Operating Information:

dBm reference = 50 Ω

Range boundaries are at voltage points, dBm levels are approximate.

$$\text{dBm} = 10 \log \left( \frac{\text{Power}}{1 \text{mW}} \right); 0.22361 \text{ V across } 50 \Omega = 1 \text{ mW or } 0 \text{ dBm}$$

- Applies to Option 5730A/05 only.

**Minimum Output.....** 300 μV (-57 dBm)

**VSWR at Output Terminal.....** <1.1 Typical

**Frequency Specification .....** ±0.01 % of output

**Frequency Resolution .....** 11,999 counts to 1.1999 MHz, 10799 counts to 11.999 MHz, 3800 counts to 50 MHz

**Overload Protection .....** A short circuit on the wideband output will not result in damage.  
After settling time, normal operation is restored upon removal.

## 52120A Specifications when Operated with the 5730A

### Line Power

Voltage range .....	100 V to 240 V
Frequency .....	47 to 63 Hz
Voltage variations .....	±10 % about line voltage
Power consumption .....	< 1500 VA

### Dimensions (HxWxL)

With feet .....	192 mm x 432 mm x 645 mm (7.6 in x 17.0 in x 25.5 in)
Without feet .....	178 mm x 432 mm x 645 mm (7.0 in x 17.0 in x 25.5 in)
Weight.....	25 kg (54 lb)

### Temperature

Operating .....	5 °C to 35 °C (41 °F to 95 °F)
Calibration (tcal) .....	16 °C to 30 °C (61 °F to 86 °F)
Storage.....	0 °C to 50 °C (32 °F to 122 °F)
Transit .....	-20 °C to +60 °C (-4 °F to +140 °F) < 100 hours

**Warmup Time** ..... Twice the time since last warmed up, to a maximum of 1 hour.

### Humidity (non-condensing)

Operating .....	< 80 %, 5 °C to 31 °C (41 °F to 88 °F) ramping linearly down to 50 % at 35 °C (95 °F)
Storage.....	< 95 %, 0 to 50 °C (32 °F to 122 °F)

### Altitude

Operating .....	2,500 m (8,200 ft) maximum
Non-Operating .....	12,000 m (39,400 ft) maximum

### Shock and Vibration

MIL-PRF-28800F Class 3

**Safety** ..... EN/IEC 61010-1, 300 V CAT II, Pollution Degree 2

**Electromagnetic Environment** ..... IEC 61326-1, Industrial

**Electromagnetic Compatibility** ..... FCC Rules part 15 sub part B

Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment)

This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

**Indoor use only**..... IP20

## 52120A Electrical Performance Limits

Voltage compliance developed across inductive loads may prevent range maximum current output being achieved at higher frequencies. The appropriate maximum frequency ( $F_{max}$ ) for a given load inductance and current is given by:

$$F_{max} = \frac{4.5}{2\pi I \cdot L} \quad I = \text{Current} \\ L = \text{Total inductance}$$

The maximum frequency calculated with this equation is only approximate. Series resistance and parallel capacitance also affect the maximum achievable frequency.

**Input Common Mode Rejection**..... 80 dB @ DC decreasing linearly to 40 dB at 10 kHz

### Input Impedance

Voltage input .....	> 1 MΩ
Current input.....	10 Ω

**Maximum Output Compliance Voltage**..... 4.5 V rms (6.4 V pk), 6.4 V dc. 120 A range maximum compliance voltage decreases from 4.5 V at 1 kHz to about 3 V at 10 kHz

**DC Offset** ..... Magnetic remanence that follows abrupt changes in output current level may cause small changes to DC current offset. It is good practice to correct for offsets in DC measurements and techniques such as DC reversal measurement will result in best accuracy.

## Operated within 5730A Control Loop (all current ranges)

The current specification of the 52120A, when controlled by a single 5730A, applies to the parallel output of up to three 52120As connected as slaves.

### Coverage factor k=2.58 (99 % confidence level)

#### Current Specification

Frequency	1-year, tcal <sup>[1]</sup> ±5 °C ±(% of output + % of range)	
	5730A	
	% of output	% of range
DC	0.015	0.010
10 Hz to 850 Hz	0.011	0.003
850 Hz to 6 kHz	0.052	0.005
6 kHz to 10 kHz	See Operated Stand Alone current specification table in the 52120A Users Manual.	

Notes:

- tcal is the temperature at which calibration adjustment took place.

Maximum inductance for stability LCOMP OFF is 100 µH. Maximum inductance for stability LCOMP ON is 400 µH for 2 A and 20 A ranges. 100 µH on the 120 A range.

With LCOMP ON, the output is limited to 7.2e3 A-Hz. For example, a 100 A output is limited to 72 Hz.

### Coverage factor k=2.00 (95 % confidence level)

#### Current Specification

Frequency	1-year, tcal <sup>[1]</sup> ±5 °C ±(% of output + % of range)	
	5730A	
	% of output	% of range
DC	0.012	0.008
10 Hz to 850 Hz	0.009	0.002
850 Hz to 6 kHz	0.040	0.004
6 kHz to 10 kHz	See Operated Stand Alone current specification table in the 52120A Users Manual.	

Notes:

- tcal is the temperature at which calibration adjustment took place.

Maximum inductance for stability LCOMP OFF is 100 µH. Maximum inductance for stability LCOMP ON is 400 µH for 2 A and 20 A ranges. 100 µH on the 120 A range.

With LCOMP ON, the output is limited to 7.2e3 A-Hz. For example, a 100 A output is limited to 72 Hz.

#### Maximum Distortion and Noise

Frequency	Distortion <sup>[1]</sup>				Noise 16 Hz to 10 MHz	
	LCOMP OFF		LCOMP ON			
	dBc	Current	dBc	Current		
<b>2 Amp Range</b>						
16 Hz to 850 Hz	-76	42 µA	-70	83 µA	-60 dB	
850 Hz to 6 kHz	-52	662 µA	-46	1.3 mA	-60 dB	
6 kHz to 10 kHz <sup>[2]</sup>	-40	2.6 mA	-35	4.7 mA	-60 dB	
<b>20 Amp Range</b>						
16 Hz to 850 Hz	-76	418 µA	-60	2.6 mA	-70 dB	
850 Hz to 6 kHz	-52	6.6 mA	-42	20.9 mA	-70 dB	
6 kHz to 10 kHz <sup>[2]</sup>	-40	26.4 mA	-35	46.9 mA	-70 dB	
<b>120 Amp Range</b>						
16 Hz to 850 Hz	-76	2.5 mA	-60	15.8 mA	-70 dB	
850 Hz to 6 kHz	-52	39.7 mA	-42	125.7 mA	-70 dB	
6 kHz to 10 kHz <sup>[2]</sup>	-40	158.2 mA	-35	281.3 mA	-70 dB	

Notes:

- Use dB or Current. Whichever is larger.
- Interharmonics only above 6 kHz.

**52120A/COIL 3 kA 25-Turn Coil****Number of Turns.....** 25**Minimum internal jaw dimension to clear wires** 26 mm (width) x 36 mm (length)**Maximum Input Current .....** 120 A continuous with built-in 12 V fan on**Maximum Voltage.....** 4.5 V rms**Specification**

Input Current <sup>[1]</sup>	Frequency	Effective Current Amp-turns	52120A + Coil Specification <sup>[2]</sup> ±(% of Amp-turns + % of 52120A range)	
			% of Amp-turns	% of 52120A Range
0 A to 100 A	DC	0 to 2500	0.7 %	0.7 %
0 A to 120 A	10 Hz to 65 Hz	0 to 3000	0.7 %	0.7 %
0 A to 120 A	65 Hz to 300 Hz	0 to 3000	0.7 %	0.7 %
0 A to 40 A	300 Hz to 1 kHz	0 to 1000	0.7 %	0.7 %
0 A to 12 A	1 kHz to 3 kHz	0 to 300	0.8 %	1.0 %
0 A to 3 A	3 kHz to 6 kHz	0 to 75	1.5 %	1.0 %
0 A to 1 A	6 kHz to 10 kHz	0 to 25	5.0 %	1.0 %

## Notes:

1. The inductance and mutual inductance of the 25 turn coil and clamp that is measured causes a frequency dependent compliance voltage across the coil. The length and configuration of the cables that connect the current to the coil also have an effect. Maximum input current is 120 A input at approximately 100 Hz. Maximum current input decreases to approximately 0.8 A at 10 kHz.
2. Includes coil/clamp interaction.

**52120A/COIL 6 kA 50-Turn Coil****Number of Turns.....** 50**Minimum Flexible Probe Length.....** 500 mm**Maximum Input Current .....** 120 A continuous with built-in 12 V fan on**Maximum Voltage.....** 4.5 V rms**Specification**

Input Current <sup>[1]</sup>	Frequency	Effective Current Amp-turns	52120A + Coil Specification <sup>[2]</sup> ±(% of Amp-turns + % of 52120A range)	
			% of Amp-turns	% of 52120A Range
0 A to 100 A	DC	0 to 5000	0.7 %	0.7 %
0 A to 120 A	10 Hz to 65 Hz	0 to 6000	0.7 %	0.7 %
0 A to 120 A	65 Hz to 300 Hz	0 to 6000	0.7 %	0.7 %
0 A to 120 A	300 Hz to 1 kHz	0 to 6000	0.7 %	0.7 %
0 A to 120 A	1 kHz to 3 kHz	0 to 6000	0.8 %	1.0 %
0 A to 25 A	3 kHz to 6 kHz	0 to 1250	1.5 %	1.0 %
0 A to 13 A	6 kHz to 10 kHz	0 to 650	5.0 %	1.0 %

## Notes:

1. The inductance and mutual inductance of the 50 turn coil causes a frequency dependent compliance voltage across the coil. Maximum frequency for 120 A input current is approximately 600 Hz. Maximum current input decreases to approximately 13 A at 10 kHz.
2. Includes coil/probe interaction.

**Note**

The specifications for these coils are at 99 % confidence level and are the combined specification of the coil and a 52120A. If the coils are used with other current sources the calibration specification of the coils alone is 0.65 % (99 % confidence level) from 0 Hz to 10 kHz.

**Operating Limits**

	Output Current Range		
	2 A	20 A	120 A
Current Output (Max.)	2 A rms	20 A rms	120 A rms
<b>Current Input</b>			
Input Current (Max.)	200 mA rms	200 mA rms	120 mA rms
Current gain	10	100	1,000
<b>Voltage Input</b>			
Input Voltage (Max.)	2 V rms	2 V rms	1.2 V rms
Transconductance	1 Siemen	10 Siemens	100 Siemens

**120 A Range Current/Frequency Limits**

Frequency	Maximum Output Current	Maximum Current Input	Maximum Voltage Input
DC	±100 A	±100 mA	±1.0 V
< 10 Hz	100 A pk (70 A rms)	100 mA pk (70 mA rms)	1.0 V pk (0.7 V rms)
10 Hz to 10 kHz	170 A pk (120 A rms)	170 mA pk (120 mA rms)	1.7 V pk (1.2 V rms)

Note:

The 2 A and 20 A ranges operate at full output current from DC to 10 kHz.

**Output Isolation**

Frequency	Maximum Voltage Signal Applied to any Output Current Terminal with respect to Earth
DC to 850 Hz	600 V rms, 850 V pk, limited 2 A rms, no transient overvoltages
850 Hz to 3 kHz	100 V rms, 142 V pk, limited 2 A rms, no transient overvoltages
3 kHz to 10 kHz	33 V rms, 47 V pk, limited 2 A rms, no transient overvoltages

## Ordering Information

### Model

5730A	Multifunction Calibrator
5730A/03	Multifunction Calibrator with 30 MHz Wideband AC Voltage Option
5730A/05	Multifunction Calibrator with 50 MHz Wideband AC Voltage Option
5730A/S	Multifunction Calibrator with no front panel USB port

### Accessories

52120A	Transconductance Amplifier
5725A	Amplifier
5730A-7002	Low Thermal EMF Cables with Banana Plugs
5730A-7003	Low Thermal EMF Cables with Spade Connectors
Y5737	5730A Rack Mount Kit
Y5738	5730A Rack Ear Kit
57XX/CASE	5730A Durable Travel Case

### Additional standards

5790B	AC Measurement Standard
734A	Voltage Reference and DC Voltage Standard
732B	Direct Voltage Standard
742A	Resistance Standards
792A	AC/DC Transfer Standard

### Software

MET/CAL	MET/CAL Plus Calibration Management Software
MET/TEAM	MET/TEAM Test Equipment Asset Management Software

**Fluke Calibration.** Precision, performance, confidence.™

▼	Electrical	RF	Temperature	Pressure	Flow	Software
---	------------	----	-------------	----------	------	----------

**Fluke Calibration**  
PO Box 9090,  
Everett, WA 98206 U.S.A.

**Fluke Europe B.V.**  
PO Box 1186, 5602 BD  
Eindhoven, The Netherlands  
Web access: <http://www.flukecal.en>

**For more information call:**

In the U.S.A. (877) 355-3225 or Fax (425) 446-5116  
In Europe/M-East/Africa +31 (0) 40 2675 200 or Fax +31 (0) 40 2675 222  
In Canada (800)-36-FLUKE or Fax (905) 890-6866  
From other countries +1 (425) 446-5500 or Fax +1 (425) 446-5116  
Web access: <http://www.flukecal.com>

©2015 Fluke Calibration. Specifications subject to change without notice.  
Printed in U.S.A. 6/2015 4265329b\_EN  
Pub-ID: 12117-eng

Modification of this document is not permitted without written permission  
from Fluke Calibration.